



**WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: June 28, 2006 INITIALS: Dec OWNER: WSP  
 DISTRICT: 2 ADDRESS: \_\_\_\_\_  
 LOCATION: 7 M Shop COUNTY: \_\_\_\_\_  
 MAKE: Hanni TYPE: port wheel weigher ACCURACY CLASS: III  
 MODEL: 101 ID. NO: 19542 CAPACITY: 20000 SECTIONS: 1  
 MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
<u>"</u>	<u>10000</u>	<u>10050</u>	<u>+50</u>				
<u>"</u>	<u>15000</u>	<u>15000</u>	<u>0</u>				
<u>Center</u>	<u>20000</u>	<u>19950</u>	<u>-50</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David E. Croner Date: 8/3/2006



State of Washington  
 County of THURSTON  
 Signed and sworn to (or affirmed) before  
 me on 8-3-2006 by  
Sylvia S. Lewallen  
 (Signature of notary public)  
SYLVIA S. LEWALLEN  
 Printed Name of Notary  
 My appointment expires 3-28-2007

3000-362-021 (R) (10/98)